

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 06 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130121</u> ✓	
Name <u>John Pernell</u>		ID# <u>6260</u> ✓	Date <u>6/28/10</u> ✓
A	Agency <u>Fairbanks Airport Police</u>		Phone # <u>907-474-2555</u>
Instrument Location <u>5195 Brumbaugh Ave. Fairbanks, AK 99709</u>			
Alco S/N <u>X124670</u> ✓		Target Value <u>.082</u> ✓	High Pressure <u>800</u>
B	Alco Test Values		
	<u>.079</u> ✓ 1 st Alco	<u>.080</u> ✓ 2 nd Alco	
Signature <u>John Pernell</u>			
(OVER)			

BMB
7/12/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 20th day of July, 2010.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130121

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130121

JUNE 28, 2010

OPERATOR'S NAME:

PERNELL/JOHN

OPERATOR'S NUMBER: 6260

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/NI :

VOC

O.L. #: VOC

DEPT/AGENCY: FBK6

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .002

ALCO S/N: X124670

— BREATH ANALYSIS —

.002 ADJUSTED FOR 29.78 in
ALCO TARGET .001 16:05
BLANK TEST .000 16:06
INTERNAL STANDARD VERIFIED 16:06
ALCO TV 29.78 in .079 16:06
BLANK TEST .000 16:07
SUBJECT SAMPLE .000 16:08
BLANK TEST .000 16:08
ALCO TV 29.78 in .000 16:09
BLANK TEST .000 16:10

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130121

JUNE 28, 2010

TIME 16:01

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 03/29/00
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 44c
BAROMETER: 29.78 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~